

GOVT OF NCT OF DELHI
NEHRU HOMOEOPATHIC MEDICAL COLLEGE & HOSPITAL
B BLOCK, DEFENCE COLONY, NEW DELHI-110024

F.N.9 (3)/NHMC/2015-16 **920**

Dated:- **07-3-2020**

To,

✓ Official website NHMC&H

Subject- Quotation for making Lead/ Lead equivalence Cabin for X-ray Department in NHMC&H.

Sir/Madam,

Sealed quotations are invited for making Lead/ Lead equivalence Cabin for X-ray Department as per below specification in Nehru Homeopathic Medical College & Hospital, B Block, Defence Colony, New Delhi-110024.

The Quotation envelopes may be marked as- Quotation for Lead/ Lead equivalence cabin.

The Specification are as under:-

Length	6 Feet
Width	7.5 Feet
Height	Approx 8 Feet upto the False Ceiling
Door	Size 7.75 X 3 Feet, should be sliding type on longer wall.
Size of Viewing Glass	14 X 14 inches of 2mm lead or lead equivalence. The same should be placed on the wall facing the patient side for direct visualization.
others	All walls of the cabin being constructed including door & window should have lead or lead equivalence of at least 2mm.
Quality Assurance	Cabin must be in compliance w.r.t. AERB radiation standards
Additional Terms condition (ATC)	1. The company must provide a certificate from an AERB authorized agency regarding the compliance of the cabin with respect to AERB radiation standards. 2. Company /firm must provide a satisfactory work compliance certificate from at least two hospital during last three years.


The sealed quotations should reach upto **16/03/2020** till 12.00 Noon, by post or be dropped in the sealed tender box placed in front of the Principal office. The tender shall be opened on the same day at 2.00 P.M.

The Items will have to be supplied at Nehru Homoeopathic Medical College & Hospital, B-Block, Defence Colony, New Delhi-110024.

The Department reserves the right to accept the quotation in part or full or reject the whole without assigning any reason. The rate should be valid till 31/03/2020.

H The Quotation is not ⁱⁿ conformity with the above specifications and ^{it} ATC will be summarily rejected. The firms interested in participating in the BID may visit the department on any working day between 10am to 3pm (Mon to Friday)

Thanking You


Dr. George Mathew
(I/C Medical Store Officer)

Copy to:- I/C Website with the request for uploading the tender on website